


Saliva DNA Storage Tube Instructions for Use

Basic Information



Jiangsu CoWin Biotech Co., Ltd.
4F, Building TQB No.1, China Medicine City Avenue, Taizhou, Jiangsu,
China
Manufacturing address:
1F and 2F No. 58, 4F East Side of No. 59, West Side of Koutai Road,
China Medicine City, Taizhou City, Jiangsu Province, China
Post code: 225300
Tel: +86-523-86201352 Fax: +86-523-86816890
Website: www.cwbiosciences.com
After-sales service Tel: +86-10-56953015
E-mail: info@cwbio.com



CMC Medical Devices & Drugs S.L
Address: C/Horacio Lengo N° 18 CP 29006, Málaga-Spain
Tel: +34 951 214054 Email: info@cmcmedicaldevices.com

Packaging Specifications

20 sets

Intended Use

The Saliva DNA Storage Tube is used for collecting and stabilizing DNA from human saliva and for the transportation and long-term room temperature storage of a saliva sample.

Saliva collected and stored in Saliva DNA Storage Tube are compatible with most commercially available nucleic acid extraction kits. And extracted nucleic acid can be for downstream analysis.

The Saliva DNA Storage Tube is intended for use by qualified clinical personnel and *in vitro* diagnostic procedures.

Components

Each set of this product is composed of 1 of Collection Tube with funnel and 1 of Saliva DNA Storage Tube with 1 mL of DNA storage buffer.

Storage Condition and Shelf Life

This product can be stored and transported below 40°C for three years prior to use.


Specimen Collection, Handling, and Storage

Saliva samples collected using the Saliva DNA Storage Tube are stabilized and can be preserved below 40°C for two years.


Saliva Collection Preparation

1. Clean the mouth with drinking water 30 minutes before saliva sample collection.
2. Relax your cheeks and gently massage cheeks with fingers for 15-30 seconds to produce saliva before collecting saliva.


Saliva Collection




1. Open package containing Collection Tube with funnel and Saliva DNA Storage Tube.




2. Spit saliva into funnel and until saliva liquid (not including bubbles) reaches fill line.



3. Hold the Collection Tube with funnel vertically, unscrew the screwcap of Saliva DNA Storage Tube and pour contents into Collection Tube with sample. Keep screwcap for Step 4.



4. Hold the Collection Tube with funnel vertically, remove funnel from Collection Tube and close with screwcap.



5. Gently invert the Saliva DNA Storage Tube 5 times to fully mix saliva with storage buffer.

6. Discard the funnel and Saliva DNA Storage Tube as requirement of local regulation.

Warnings and Precautions

1. For *in vitro* diagnostic use (IVD) only.
2. Check the package before use. If damaged, it is strictly prohibited to use.
3. DO NOT drink, touch or remove the reagent from the tube.
4. Avoid reuse.
5. If any serious incident that has occurred in relation to the product shall be reported to the manufacturer and the competent authority.
6. Do not eat, drink, smoke, use tobacco products, or chew gum for at least 30 minutes before giving your saliva sample.
7. If the contents of the tube contact your skin, flush thoroughly with water.
8. Most people take 2-5 minutes to collect saliva sample following the steps above.



GHS07

H302 Harmful if swallowed.
H315 Causes skin irritation.
H319 Causes serious eye irritation.

Glossary of Symbols

Symbol	Meaning	Symbol	Meaning
	<i>In vitro</i> diagnostic medical device		Temperature limit
	Manufacturer		Use-by date
	Date of Manufacture		Do not re-use
	Batch code		Consult instructions for use
	Authorized representative in the European Community / European Union		CE Symbol
	Caution		Unique device identifier
	Catalogue number		Do not use if package is damaged and consult instructions for use